Instruction Manual

Washing and Autopsy Table

Type: WT 10/670

KUGEL medical GmbH & Co. KG
Hermann-Köhl-Str. 2A
93049 Regensburg / GERMANY

Phone +49 941/20 86 48-0
Fax +49 941/20 86 48-29
www.KUGEL-medical.de

zertifiziert nach ISO 9001

KUGEL medical

Instruction Manual

Content

1 General Information 3
1.1 Symbols in this Instruction Manual 3
1.2 Indications for Use 3
1.3 General Information and Maintenance 4

2 General Maintenance Tips for Stainless Steel Surfaces 4
2.1 Initial Cleaning / Cleaning / Preventive Measures 4
2.2 Cleaning Agents / Cleaning Instruments 5
2.2.1 Stainless Steel Polish / Stainless Steel Cleaner 5
2.2.2 Suitable Cleaning Agents 5
2.2.3 Descaling Agents 5
2.2.4 Brushes / Non-Woven Abrasives 5
2.2.5 Textiles 6

3 Declaration of Conformity 7

KUGEL medical GmbH & Co. KG
Hermann-Köhl-Str. 2A
93049 Regensburg / GERMANY

Phone +49 941/20 86 48-0
Fax +49 941/20 86 48-29
www.KUGEL-medical.de

zertifiziert nach ISO 9001

KUGEL medical

Instruction Manual

1 General Information
1.1 Symbols in this Instruction Manual

Warning, Caution
Failure to follow these instructions can lead to personal injuries, property damages and, at worst, death.

☞ Additional information

1.2 Indications for Use

Press the fasteners and push against the table legs at the same time at one site of the table.

Caution: Risk of Danger! Push against the table legs until you hear a „click“

Hold the washing table at the table plate and fold in the other table legs.

Caution: Risk of Danger! Push against the table legs until you hear a „click“

Now check if the table legs are folded in correctly otherwise the table legs can suddenly unfold which might lead to severe injuries.

Repeat the above-mentioned steps in revised order to unfold the table.

1.3 General Information and Maintenance

The mobile washing and autopsy table is absolutely maintenance-free. Maintenance is limited to following points:

- Make sure that all table legs have full ground contact to avoid annoying shaking if the floor is uneven.
- Regular cleaning of complete table
- Regular testing of all fasteners and table legs

It is forbidden to use the washing and autopsy table for any other purpose. The operating permits and all warranty claims will be forfeited if this rule is violated.

2 General Maintenance Tips for Stainless Steel Surfaces

The best thing about stainless steel is that it barely requires maintenance and soiling can be removed easily.

2.1 Initial Cleaning / Cleaning / Preventive Measures

After installing the stainless steel devices and/or furniture, an initial cleaning is to be performed to remove any soiling caused by storage, transport or installation works.

Protective films, paper packaging or residues of adhesives are to be removed completely after having installed the devices or furniture otherwise they might lead to corrosion. For cleaning you may only use suitable cleaning agents (see below). Normally, the corresponding cleaning agent is being distributed evenly on the surface with a microfiber cloth. Tough soiling is to be treated with a soft synthetic fleece, e.g.

Then remove cleaning agent stains with clear water and dry the surfaces with soft cloths.

Poorly soluble residues or small scratches can be removed with a suitable non-woven abrasive.

2.2 Cleaning Agents / Cleaning Instruments

Cleaning agents and instruments are t be used in accordance with the instructions of the respective producer.

2.2.1 Stainless Steel Polish / Stainless Steel Cleaner

Stainless steel polish – special cleaning product and protective agent for stainless steel. 400ml spray can.
Article No. 060.010.001

Shake spray can before use. Spray from approx. 25 cm distance and let it sink in for a few minutes. Then remove with a clean, dry cloth. You may sterilize and disinfect the surface if necessary afterwards.

2.2.2 Suitable Cleaning Agents

Basically, alkaline cleaning agents are to be used on stainless steel parts only.

Never use cleaning agents that contain hydrochloric acids or bleach. They might lead to discolorations.

2.2.3 Descaling Agents

Only use off-the shelf descaling agents for stainless steel.

2.2.4 Brushes / Non-Woven abrasives

Brushes made of natural or plastic bristles

Article	Article No.	Grain Size	Dimension L/W
Nylon non-woven abrasive (fine grain)	060.010.002	P 1000	200 x 115mm
Nylon non-woven abrasive (coarse grain)	060.010.003	P 280	200 x 115mm

Cleaning instructions for non-woven abrasives are to be observed provided that it is a ground surface. Always clean towards the grinding pattern, never across.

2.2.5 Textiles

Textiles made of natural or chemical fibers. Particularly microfibers are suitable to remove patches or fingerprints from stainless steel surfaces.

Article	Article No	Material	Dimension L/W
Microfiber cloth	060.010.005	80% Polyester 20% Polyamides	400 x 400 mm

KUGEL medical GmbH & Co. KG
Hermann-Köhl-Str. 2A
93049 Regensburg / GERMANY

Phone +49 941/20 86 48-0
Fax +49 941/20 86 48-29
www.KUGEL-medical.de

DECLARATION OF CONFORMITY

We KUGEL medical GmbH & Co. KG
Hermann-Köhl-Straße 2A
DE - 93049 Regensburg / GERMANY

explain in own responsibility, that the product:

Kind of equipment:	Washing and autopsy tables

Type:	WT 10/650, WT 10/655, WT 10/670

is in compliance with following standard specifications or documents

Technical regulations:	2006/42/EG

Regensburg, 02.05.2018

Authorized representative signature